

TENDER SCHEDULE NO:-R/4

Bill of Quantities/ Schedule of Requirements for TRADE : "WIREMAN".

Shortfall of Machinery

DELEVERY POINTS:- B.P.C. ITC, P.O. Krishnagar, Dist. Nadia, PIN :- 741101

Sl.No.	Description of Item with Specification	QtyRequired	UNIT	TOTAL AMOUNT (Rs)	
				In figures	In words
1	DC Compound 3HP 250V with 4 Point Starter and Laboratory type Motor should be mounted on a base plate. The Panel Board Consist of : <ul style="list-style-type: none">• 4 Point DC Starter.• Field Regulator• Suitable Analog Voltmeter.• Suitable Analog Ammeter. Accessories: • Tachometer for Speed Measurement. • The circuit diagram will be supplied with the Motor. • Instruction Manual	1 No.	each		
2	Universal Motor 750W AC/DC 250V Laboratory type Motor should be mounted on a base plate. The Panel Board Consist of : <ul style="list-style-type: none">• Suitable Analog Voltmeter.• Suitable Analog Ammeter. Accessories: • Tachometer for Speed Measurement. • The circuit diagram will be supplied with the Motor. • Instruction Manual	1 No.	each		
3	Oscillator AF Specification : Switch : 10 Pre selected spot frequencies Selectable : in 20Hz to 20KHz band Spot Frequencies : 20Hz, 100Hz, 1KHz, 2KHz, 5KHz, 8KHz, 10KHz, 14KHz, 17KHz & 20KHz. Continuously Variable : 30Hz – 300Hz & 300Hz -3KHz Output Level : (a) Output 1 : Max. 10V rms (b) Output 2 : Max 270 rms Frequency Response : ± 2 dB Harmonic Distortion : Within 2%	4 Nos.	each		

	Output Power (a) Output 1 : Max. 10W to a 10E load (b) Output 1 : 7W to a 10K load.				
4	Oil Testing Kit Confirms to IS 6792, IEC 156 Tests insulating oils to IS, IEC, ASTM, etc. Test voltage to 0 – 60 kV. One piece construction with safety interlock Test voltage can be held at any level up to maximum kV meter continues to indicate 'bdv' even after breakdown Light weight, portable, rugged and reliable Specifications : Input : 240 V \pm 10%, AC, 1Phase, 50/60Hz. Display : Analog models: 96mm square Accuracy : \pm 4% of full scale deflection/range Shut down sensitivity in HV circuit: 20mA approx. Rate of rise of voltage : Motorized sets only: 2kV/second approx. (\pm 20% max.) Dimension : 50x55x48	2 Nos.	each		

We agree to supply the above goods in accordance with the technical specification as per terms and conditions specified in Tender Notice.

We also confirm that the normal commercial Warrantee/Gurantee of 12 months shall apply to all the offered goods.

We also confirm that we have taken steps to ensure that no person acting for us or on our behalf will engage in corrupt, fraudulent, collusive, coercive or obstructive practice.

We agree to make necessary arrangements for Inspection and Testing of all the above items at places within Kolkata Metropolitan area before dispatch to the consignee.

Signature of Supplier